

CERTIFICATE OF CONFORMITY



A D T

Equipment : Logic Analyzer
Brand Name : ZEROPLUS
Test Model No. : LAP-F2
Multiple Listing : LAP-FN (N: 0-9, A-Z)
Applicant : Zeroplus Technology Co., Ltd.
Test Report No. : CE150504D02 R2



We, **Bureau Veritas Consumer Products Services (H.K.) Ltd., Taoyuan Branch**, declare that the equipment above has been tested in our facility and found compliance with the requirement limits of applicable standards, in accordance with the Directives 2004/108/EC. The test record, data evaluation and Equipment Under Test (EUT) configurations represented herein are true and accurate under the standards herein specified.

EN 61326-1:2013

EN 61326-2-2:2013

EN 61000-3-2:2014

EN 61000-3-3:2013

EN 55011:2009 +A1:2010, Group I, Class A

EN 61000-4-2:2009/ IEC 61000-4-2:2008 ED. 2.0

EN 61000-4-3:2006 +A1:2008 +A2:2010/ IEC 61000-4-3:2010 ED. 3.2

EN 61000-4-4:2012/ IEC 61000-4-4:2012 ED. 3.0

EN 61000-4-5:2006/ IEC 61000-4-5:2005 ED. 2.0

EN 61000-4-6:2014/ IEC 61000-4-6:2013 ED. 4.0

EN 61000-4-8:2010/ IEC 61000-4-8:2009 ED. 2.0

EN 61000-4-11:2004/ IEC 61000-4-11:2004 ED. 2.0

NOTE: The above EN/IEC basic standards are applied with latest version if customer has no special requirement.

Henry Lai / Director

Nov. 9, 2015



No. 47-2, 14th Ling, Chia Pau Vil., Lin Kou Dist., New Taipei City, Taiwan (R.O.C.)

Tel: 886-2-26052180 Fax: 886-2-26051924

<http://www.bureauveritas-adt.com> E-Mail: service.adt@tw.bureauveritas.com

Cancels and replaces the certificate no.: CE150504D02 R1C dated Nov. 9, 2015